

# Instructions for Use

## Disposable Forceps (VS07XX.2X-XX)

**Indications for use** - Vortex Surgical Disposable Forceps are intended as a single-use instrument used in ophthalmic surgery for posterior segment procedures.

**Contraindications** - None

**Known complications** - Retinal detachment, tears, holes and contusions, bleeding, inflammation, and infection.

These complications are statistically rare and it is assumed that the user is adequately trained in the treatment of these known complications and methods of avoidance.

**Caution** - Federal (USA) law restricts this device sale by, or on the order of, a physician.

### Instructions for use:

1. Determine package integrity.
2. Remove product from protective packaging.
3. Verify tip operation by actuating the handle and determine if tip is damaged. Do not attempt to repair or modify instrument.
4. Affect sterile transfer of the product to the sterile field.
5. Follow established surgical procedures.

**Device Storage:** Vortex Surgical, Inc recommends that the product is stored in a clean, dry and well-ventilated area at room temperature 15-37 °C (59-98.6 °F) away from direct sunlight.

**Sterilization:** This product is sterilized by Ethylene Oxide Gas (EtO) and provided with an EtO indicator. Verify indicator status color prior to use.

### Warnings:

- Do not use if package integrity has been compromised.
- Do not use if tip or handle is damaged.
- Do not use if product exposed to conditions outside of indicated range.
- Do not use if EtO indicator does not show exposure to sterilization gas.

**Reuse:** Do not reuse or reprocess instrument.

**Re-sterilization:** Do not re-sterilize instrument

**Device Disposal:** This single use surgical instrument should be considered clinical waste and should be disposed in accordance with clinical waste laws applicable in your country.


**Reporting:** Report serious incidents to Vortex Surgical and your regulatory body

Vortex Surgical, Inc. excludes all warranties, whether expressed, implied, or otherwise, to matters beyond the direct control of Vortex Surgical, Inc and the end result of this device's use. This would include, but not be limited to handling, shipment and storage of the device and patient diagnosis and treatment. The fitness and merchantability of this device are as specified. Implied factors are specifically excluded. Vortex Surgical, Inc. is not liable for loss, whether incidental or consequential, damage and/or expense, arising directly or indirectly from the use of this device. There is no additional liability or responsibility assumed, other than that specified. All additional liability or responsibility is specifically disclaimed.

Symbol glossary available on website ([www.vortexsurgical.com/symbols](http://www.vortexsurgical.com/symbols))




Vortex Surgical, Inc. 4 Research Park, Suite 124, Saint Charles, MO 63304, USA  
info@vortexsurgical.com 636-778-4350 www.vortexsurgical.com




© Vortex Surgical Inc., 2022

2200003 rev J



© Vortex Surgical Inc., 2022

2200003 rev J



© Vortex Surgical Inc., 2022

2200003 rev J

# Instructions for Use

## Disposable Forceps (VS07XX.2X-XX)

**Indications for use** - Vortex Surgical Disposable Forceps are intended as a single-use instrument used in ophthalmic surgery for posterior segment procedures.

**Contraindications** - None

**Known complications** - Retinal detachment, tears, holes and contusions, bleeding, inflammation, and infection.

These complications are statistically rare and it is assumed that the user is adequately trained in the treatment of these known complications and methods of avoidance.

**Caution** - Federal (USA) law restricts this device sale by, or on the order of, a physician.

### Instructions for use:

1. Determine package integrity.
2. Remove product from protective packaging.
3. Verify tip operation by actuating the handle and determine if tip is damaged. Do not attempt to repair or modify instrument.
4. Affect sterile transfer of the product to the sterile field.
5. Follow established surgical procedures.

**Device Storage:** Vortex Surgical, Inc recommends that the product is stored in a clean, dry and well-ventilated area at room temperature 15-37 °C (59-98.6 °F) away from direct sunlight.

**Sterilization:** This product is sterilized by Ethylene Oxide Gas (EtO) and provided with an EtO indicator. Verify indicator status color prior to use.

### Warnings:

- Do not use if package integrity has been compromised.
- Do not use if tip or handle is damaged.
- Do not use if product exposed to conditions outside of indicated range.
- Do not use if EtO indicator does not show exposure to sterilization gas.

**Reuse:** Do not reuse or reprocess instrument.

**Re-sterilization:** Do not re-sterilize instrument

**Device Disposal:** This single use surgical instrument should be considered clinical waste and should be disposed in accordance with clinical waste laws applicable in your country.

**Reporting:** Report serious incidents to Vortex Surgical and your regulatory body

Vortex Surgical, Inc. excludes all warranties, whether expressed, implied, or otherwise, to matters beyond the direct control of Vortex Surgical, Inc and the end result of this device's use. This would include, but not be limited to handling, shipment and storage of the device and patient diagnosis and treatment. The fitness and merchantability of this device are as specified. Implied factors are specifically excluded. Vortex Surgical, Inc. is not liable for loss, whether incidental or consequential, damage and/or expense, arising directly or indirectly from the use of this device. There is no additional liability or responsibility assumed, other than that specified. All additional liability or responsibility is specifically disclaimed.

Symbol glossary available on website ([www.vortexsurgical.com/symbols](http://www.vortexsurgical.com/symbols))



Vortex Surgical, Inc. 4 Research Park, Suite 124, Saint Charles, MO 63304, USA  
info@vortexsurgical.com 636-778-4350 www.vortexsurgical.com



Vortex Surgical, Inc. 4 Research Park, Suite 124, Saint Charles, MO 63304, USA  
info@vortexsurgical.com 636-778-4350 www.vortexsurgical.com



# Instructions for Use

## Disposable Forceps (VS07XX.2X-XX)

**Indications for use** - Vortex Surgical Disposable Forceps are intended as a single-use instrument used in ophthalmic surgery for posterior segment procedures.

**Contraindications** - None

**Known complications** - Retinal detachment, tears, holes and contusions, bleeding, inflammation, and infection.

These complications are statistically rare and it is assumed that the user is adequately trained in the treatment of these known complications and methods of avoidance.

**Caution** - Federal (USA) law restricts this device sale by, or on the order of, a physician.

### Instructions for use:

1. Determine package integrity.
2. Remove product from protective packaging.
3. Verify tip operation by actuating the handle and determine if tip is damaged. Do not attempt to repair or modify instrument.
4. Affect sterile transfer of the product to the sterile field.
5. Follow established surgical procedures.

**Device Storage:** Vortex Surgical, Inc recommends that the product is stored in a clean, dry and well-ventilated area at room temperature 15-37 °C (59-98.6 °F) away from direct sunlight.

**Sterilization:** This product is sterilized by Ethylene Oxide Gas (EtO) and provided with an EtO indicator. Verify indicator status color prior to use.

### Warnings:

- Do not use if package integrity has been compromised.
- Do not use if tip or handle is damaged.
- Do not use if product exposed to conditions outside of indicated range.
- Do not use if EtO indicator does not show exposure to sterilization gas.

**Reuse:** Do not reuse or reprocess instrument.

**Re-sterilization:** Do not re-sterilize instrument

**Device Disposal:** This single use surgical instrument should be considered clinical waste and should be disposed in accordance with clinical waste laws applicable in your country.

**Reporting:** Report serious incidents to Vortex Surgical and your regulatory body

Vortex Surgical, Inc. excludes all warranties, whether expressed, implied, or otherwise, to matters beyond the direct control of Vortex Surgical, Inc and the end result of this device's use. This would include, but not be limited to handling, shipment and storage of the device and patient diagnosis and treatment. The fitness and merchantability of this device are as specified. Implied factors are specifically excluded. Vortex Surgical, Inc. is not liable for loss, whether incidental or consequential, damage and/or expense, arising directly or indirectly from the use of this device. There is no additional liability or responsibility assumed, other than that specified. All additional liability or responsibility is specifically disclaimed.

Symbol glossary available on website ([www.vortexsurgical.com/symbols](http://www.vortexsurgical.com/symbols))